

December 2008



Innovative vision, solutions for life™

Propulse NG - Maintenance & Safety Inspections

To ensure optimum performance the Propulse NG Electronic Ear Irrigator should be serviced every 12 months. Service or repairs conducted by unauthorised agencies/organisations invalidate any or implied warranties from Mirage.

The Propulse NG should under-go routine electrical safety testing to ensure that it remains safe to use.

Users of the Propulse NG should carry out regular inspections to ensure that the handle and hose, power adaptor and cable, reservoir and main body of the unit are free from damage prior to use. If any damage is evident the Propulse NG should **NOT** be used until replacement parts have been fitted. Only Propulse branded items should be used with the Propulse NG.

The Propulse NG is not user repairable and should be returned to your Propulse supplier or Mirage Service Department (UK only) for service/and or repair:

Mirage Service Department
53 Purbeck Drive, Lostock, Bolton,
Lancashire, BL6 4JF

Office: +44 (0) 845 130 5440
Fax: +44 (0) 845 130 6440

BioPark Hertfordshire
Broadwater Road
Welwyn Garden City
Hertfordshire AL7 3AX UK

info@miragehealthgroup.com
www.miragehealthgroup.com



MD 523880